



RE: CLAVE® Connector in Nuclear Medicine

Thank you for your interest in the CLAVE Connector for use in nuclear medicine. The deadspace of the CLAVE connector is 0.06cc which results in minimal hold-up volume of the Isotope. The primary concern when administering Isotopes is the 'loose' contamination which can shed into the environment. ICU recommends briskly swabbing the CLAVE following an Isotope infusion to remove the loose particulate. The silicone seal of the CLAVE may absorb some molecules of the Isotope as most plastics do, but these are described as 'fixed' contamination which does not shed into the environment.

Thank you again for your interest, if there are further questions or concerns please visit our website at www.icumed.com, or contact the corporate offices at 949-366-2183 or 800-824-7890.

Technical Services
ICU Medical, Inc.